

**EU Declaration of Conformity**Manufacturer's Name: **AMETEK Denmark A/S**Manufacturer's Address: Gydevang 32-34  
3450 Allerød - Denmark

We declare that this Declaration of Conformity is issued under our sole responsibility and belongs to the following products:

Product Name:	JOFRA CTC-Series Temperature Calibrator
Model Number:	CTC-155 A, CTC-155 C from serial no. xxxxxx-00729 CTC-350 A, CTC-350 C from serial no. xxxxxx-00342 CTC-652 A, CTC-652 C from serial no. xxxxxx-00001 CTC-660 A, CTC-660 C from serial no. xxxxxx-00581
Product Name:	JOFRA MTC – Marine Temperature Calibrator
Model Number:	MTC-650 A MKII from serial no. xxxxxx-00056

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation:

- **EMC Directive 2014/30/EU**
  - Has demonstrated fulfilment of the essential requirements set out in Annex I.
- **Low Voltage Directive 2014/35/EU:**
  - Has demonstrated fulfilment of the safety objectives referred to in Article 3 and set out in Annex I.
- **RoHS Recast (RoHS II) Directive 2011/65/EU**
  - Has demonstrated that the requirements specified in article 4 have been met.

The following harmonized standards and technical specifications have been applied:

**EMC Directive 2014/30/EU:**  
EN 61326-1:2013Electrical equipment for measurement, control  
And laboratory use – EMC requirements  
Part 1: General requirements  
Emission : Class B - equipment suitable for use in domestic establishments and in  
establishments directly connected to low voltage power supply  
network which supplies buildings used for domestic purposes.  
  
Immunity: Table 2 – Immunity test requirements for equipment intended to be used in  
an industrial electromagnetic environment.**Low Voltage Directive 2014/35/EU:**

EN 61010-1:2010

Safety requirements for electrical equipment for Measurement, control and laboratory  
use –  
Part 1: General requirements

EN61010-2-010: 2014

Safety requirements for electrical equipment for measurement, control and laboratory  
use -  
Part 2: Particular requirements for laboratory equipment for the heating of materials  
\_\_\_\_\_  
Signed for and on behalf of AMETEK Denmark A/S

Name: Joel Frie

Title: Division Vice President & Business Unit Manager  
AMETEK Measurement, Communications & Testing Division  
AMETEK Sensors, Test & Calibration Business Unit

Place of Issue: AMETEK Denmark A/S Gydevang 32-34, 3450 Allerød, Denmark

Date of Issue: 23 January 2020