



## Declaration of Conformity

Manufacturer's Name: **AMETEK Denmark A/S**

Manufacturer's Address: Gydevang 32-34  
3450 Allerød - Denmark

We declare that this Declaration of Conformity is issued under our sole responsibility and belongs to the following products:

Product Name: CTC-Series Temperature Calibrator  
Model Number: CTC-1205 A, CTC-1205 C from serial no. xxxxxx-00157

The object of the declaration described above is in conformity with the relevant statutory requirements:

- Electromagnetic Compatibility Regulations 2016
- Electrical Equipment (Safety) Regulations 2016
- The Restriction of the Use of Certain Hazardous Substances in electrical and Electronic Equipment Regulations 2012


The following harmonized standards and technical specifications have been applied:

### Electromagnetic Compatibility Regulations 2016:

EN 61326-1:2013 Electrical equipment for measurement, control  
And laboratory use – EMC requirements  
Part 1: General requirements  
Emission: Class B - equipment suitable for use in domestic establishments and in  
establishments' directly connected to low voltage power supply  
network which supplies buildings used for domestic purposes.  
Immunity: Table 2 – Immunity test requirements for equipment intended to be used in  
an industrial electromagnetic environment.

### Electrical Equipment (Safety) Regulations 2016:

EN 61010-1:2010 Safety requirements for electrical equipment for Measurement, control and laboratory  
use –  
Part 1: General requirements  
EN61010-2-010: 2014 Safety requirements for electrical equipment for measurement, control and laboratory  
use -  
Part 2: Particular requirements for laboratory equipment for the heating of materials

  
Signed for and on behalf of AMETEK Denmark A/S

Name: Joel Frie  
Title: MCT Division Vice President & STC Business Unit Manager  
AMETEK Measurement, Communications & Testing Division  
AMETEK Sensors, Test & Calibration Business Unit  
Place of Issue: AMETEK Denmark A/S Gydevang 32-34, 3450 Allerød, Denmark  
Date of Issue: 17 September 2021