



Declaration of Conformity

Manufacturer's Name: **AMETEK Denmark A/S**

Manufacturer's Address: Gydevang 32-34
3450 Allerød - Denmark

We declare that this Declaration of Conformity is issued under our sole responsibility and belongs to the following products:

Product Name: DTI Series Thermometer
Model Number: DTI-1000 A, DTI-1000 B from serial no. xxxxxx-00649


The object of the declaration described above is in conformity with the relevant statutory requirements:

- Electromagnetic Compatibility Regulations 2016
- The Restriction of the Use of Certain Hazardous Substances in electrical and Electronic Equipment Regulations 2012

The following harmonized standards and technical specifications have been applied:

Electromagnetic Compatibility Regulations 2016:

EN 61326-1:2013 Electrical equipment for measurement, control
And laboratory use – EMC requirements
Part 1: General requirements
Emission: Class B - equipment suitable for use in domestic establishments and in
establishments' directly connected to low voltage power supply
network which supplies buildings used for domestic purposes.
Immunity: Table 1 – Immunity test requirements for equipment intended to be used in
an basic electromagnetic environment.
Accuracy can be reduced to 0.028°C in basic electromagnetic
environment.
Table 3 – Immunity test requirements for equipment intended to be used in
a controlled electromagnetic environment.


Signed for and on behalf of AMETEK Denmark A/S

Name: Joel Frie
Title: MCT Division Vice President & STC Business Unit Manager
AMETEK Measurement, Communications & Testing Division
AMETEK Sensors, Test & Calibration Business Unit
Place of Issue: AMETEK Denmark A/S Gydevang 32-34, 3450 Allerød, Denmark
Date of Issue: 17 September 2021