

## Declaration of Conformity

Manufacturer's Name: **AMETEK Denmark A/S**

Manufacturer's Address: Gydevang 32-34  
3450 Allerød - Denmark

We declare that this Declaration of Conformity is issued under our sole responsibility and belongs to the following product:

Product Name: mAcal Series Loop Calibrator  
Model Number: mAcal from serial no. xxxxxx-09033

The object of the declaration described above is in conformity with the relevant statutory requirements:

- Electromagnetic Compatibility Regulations 2016
- The Restriction of the Use of Certain Hazardous Substances in electrical and Electronic Equipment Regulations 2012

The following harmonized standards and technical specifications have been applied:

**Electromagnetic Compatibility Regulations 2016:**

EN 61326-1:2013

Electrical equipment for measurement, control  
And laboratory use – EMC requirements  
Part 1: General requirements

Emission: Class B - equipment suitable for use in domestic establishments and in establishments' directly connected to low voltage power supply network which supplies buildings used for domestic purposes.

Immunity: Table 1 – Immunity test requirements for equipment intended to be used in a basic electromagnetic environment.

Table 2 – Immunity test requirements for equipment intended to be used in an industrial electromagnetic environment.

mA-source accuracy is 0.2% FS in industrial electromagnetic environment.



Signed for and on behalf of AMETEK Denmark A/S

Name: Joel Frie

Title: MCT Division Vice President & STC Business Unit Manager  
AMETEK Measurement, Communications & Testing Division  
AMETEK Sensors, Test & Calibration Business Unit

Place of Issue: AMETEK Denmark A/S Gydevang 32-34, 3450 Allerød, Denmark

Date of Issue: 17 September 2021